



KEIYU NDT SUPPLY COMPANY

1F., No.409, Sec. 6, Zhongxiao E. Rd., Nangang Dist., Taipei City 115, Taiwan (R.O.C.)
TEL: 886-2-26534530, 26546891 FAX: 886-2-26547597
E-Mail : sale@keiyu-ndt.com.tw Website: www.keiyu-ndt.com.tw

Customer: To Whom It May Concern.

Certificate of Conformity

This is to certify that the Magnetic Field Indicator supplied to whom it may concern were inspected and qualified by KEIYU NDT SUPPLY COMPANY.

All measurements were performed at standard room temperature 23 ° C.
All gauges were tested by SGS with standards traceable to NIST, and meet the following specification:

ASME, SECTION V, ARTICLE 7
ASME, SECTION V, ARTICLE 25
MIL-STD-271F(SH), PARAGRAPH 4.3.1.1, FIGURE 6
NAVSEA Tech PUB T904-AS-GIB-010/271, PARA.4.3.1.8, Fig. 16
NAVSHIPS 250-1500-1, REV. 11, PARAGRAPH 12.4.1.8, FIGURE 12-18
MIL-STD-2132A, FIGURE 7
ASTM E709-95, PARAGRAPH 14.2.2, FIGHRE 15
MIL-STD-1949, PARAGRAPH 5.8.3, FIGURE 6
ASTM E1444-01, PARAGRAPH 7.1.2, FIGURE 3